RETURN TO TECHNICAL SERVICE BULLETIN TABLE OF CONTENTS
RETURN TO CD-ROM TABLE OF CONTENTS

TECHNICAL SERVICE BULLETIN #6

Product: Fabius GS

Date: April 26, 2003

Re: Software Release - 1.37

Reference Doc: Product Bulletin # 149

Reason/Rational: Enhancement

Solution/Procedure: If applicable, update 1.35 (field) Software kits to 1.37. Perform Software 1.37 upgrade on Fabius GS units, complete the Return Form (attached) and return to DMI/DS within 7 days of account completion.

Devices affected: All

Timing/When: At customer request, next service or Periodic Manufacturer's Service

Additional Info: Software 1.37 supersedes software 1.35 upon release. Therefore, all current field activity to upgrade units to 1.35, shall proceed with 1.37 instead. Current 1.35 software kits located in the field shall be updated to 1.37 by ordering P/N 4117937-004 - Software 1.37 CD.

Software upgrades from 1.3n to 1.37 shall be completed using Software Kit - 4118364

Note: Units with software 1.1n or 1.20 to 1.3x shall only be performed using the appropriate upgrade kit, which includes the necessary hardware**.** Refer to Product Bulletin # 149 for details.

Cost: Software upgrades to 1.37 are available to Fabius GS customers at no charge.

Ordering Info/Parts: Contact your DMI Customer Service Representative to order all software kits.

Distribution: DrägerService & All Authorized Service Organizations worldwide.

DrägerService
3122 Commerce Drive
Telford, PA 18969
Tel: (215) 721-5402
(800) 543-5047
Fax: (215) 723-5935
Web: www.draegermedical.com
DrägerService is a division of
Draeger Medical, Inc.

RETURN TO TECHNICAL SERVICE BULLETIN TABLE OF CONTENTS
RETURN TO CD-ROM TABLE OF CONTENTS

Return Form - Software 1.37 Upgrade

Distributor:	Customer:
Country:	Address:
	City:

Machine S/N	Machine S/N	Machine S/N
_____	_____	_____
_____	_____	_____
_____	_____	_____
_____	_____	_____
_____	_____	_____
_____	_____	_____
_____	_____	_____
_____	_____	_____
_____	_____	_____
_____	_____	_____
_____	_____	_____
_____	_____	_____
_____	_____	_____
_____	_____	_____
_____	_____	_____
_____	_____	_____
_____	_____	_____
_____	_____	_____
_____	_____	_____

Fax To:
DrägerService
Attn: Al Myers
(215) 721-5784

Date of Update _____
Signature (TSR) _____
Name (Print) _____

DrägerService
3122 Commerce Drive
Telford, PA 18969
Tel: (215) 721-5402
(800) 543-5047
Fax: (215) 723-5935
Web: www.draegermedical.com
DrägerService is a division of
Draeger Medical, Inc.